



ISSN Print: 2394-7489
ISSN Online: 2394-7497
IJADS 2023; 9(2): 506-510
© 2023 IJADS
www.oraljournal.com
Received: 02-03-2023
Accepted: 03-04-2023

Mohammed Labib
Professor of Fixed Department
of Prosthodontics, Faculty of
Dentistry, Cairo University,
Cairo, Egypt

Zamzam, Samaa Kotb
Lecturer of Fixed Department of
Prosthodontics, Faculty of
Dentistry, Cairo University,
Cairo, Egypt

Toaa Mohammed Saad
B.D.S Faculty of Dentistry,
Cairo University, Cairo, Egypt

Evaluation of marginal adaptation and internal fit of maxillary complex provisional bridge manufactured by two different methods: An *In-vitro* study

Mohammed Labib Zamzam, Samaa Kotb, Toaa Mohammed Saad

DOI: <https://doi.org/10.22271/oral.2023.v9.i2g.1766>

Abstract

Aim: The purpose of this in vitro study was to evaluate if the marginal discrepancy and internal fit of maxillary complex provisional bridge is improved by either 3D printing or CAD/CAM fabrication technique

Methodology: In the present study a total of sixteen samples (Eight in each group) fabricated by two methods using same STL file. Eight samples were fabricated from Telio CAD by Subtractive CAD CAM method (ARUM 5X- 400 Milling machine) (Control group), the other Eight samples were fabricated from HARZ Labs Dental Sand liquid resin by additive 3D printing (Anycubic photon S printer) (intervention group). All samples were tested for marginal gap by Digital microscope and internal fit was tested by silicon replica technique.

Results: It was found that regarding marginal adaptation and internal fit results; no statistically significant difference between the fabrication methods.

Conclusion: Marginal fit and internal adaptation of (Additive three dimensional printing and Subtractive milling) fabrication techniques were within clinically acceptable ranges.

Clinical implication: Both (Additive. and. subtractive) techniques could be used for provisional bridge fabrication

Keywords: 3D printing, additive, subtractive, milling, CAD/CAM, marginal fit and internal adaptation

Introduction

A dental prosthesis is used to restore oral function and aesthetics when replacing missing teeth, this allows the preservation and improvement of the patient's appearance, comfort, physical and mental health. variety of treatment methods are available for restoring missing teeth dependent on the number and state of remaining teeth, existing space, bone support, expense, and patient desire. (Abduo *et al.*, 2014) ^[1].

The standardised digital systems have efficiently substituted the conventional impressions by the optical impressions and the casting techniques were substituted by the computer aided designing (CAD) and the computer aided manufacturing (CAM). Even in a digital system, it is necessary to create a functional model in order to determine and correct the restoration's fit. In addition, the working model will allow for improved restoration planning and design, as well as the combination of conventional and digital techniques. (Dureja *et al.*, 2018) ^[6]

Three dimensional printing began to enter this field in order to solve some of the shortcomings of the milling technique. Additive manufacturing produces precise accurate prosthesis by minimal materials and less cost. Additionally, multiple restorations can be fabricated simultaneously. However, the primary disadvantage of this method of production is the co-occurring dimensional discrepancy that can manifest in the final model or restoration as various forms of clinical inaccuracy due to shrinkage during the building process and post-curing. (Jeong *et al.*, 2018) ^[9].

Material and methods

In a dentate typodont model maxillary central and lateral incisors were prepared by free hand

Corresponding Author:
Toaa Mohammed Saad
B.D.S Faculty of Dentistry,
Cairo University, Cairo, Egypt

using depth grooves, with 2mm incisal reduction, 1.5 mm axial, and finish line 1 mm circumferential deep chamfer. Maxillary first premolar prepared by free hand using depth grooves, with 1.5 mm on buccal cusp and 2mm on palatal cusp, 1.5mm axial, with a finish line 1mm deep chamfer. Through tapered; stone with rounded; end size 13 (TR 13),

size 14 (TR14), needle stone and football stone. any sharp areas that might perform as a point for stress concentration were rounded by the fine grit 852F. Rubber base index and graduated periodontal probe were used to confirm preparation thickness. Figure (1)




Fig 1: a) Finished and polished prepared abutments, b) Confirmation of preparation thickness using periodontal probe with the rubber base index.

Digital impression was taken by Medit i500 intra-oral scanner. The same STL file was used for designing both 3D

printing and subtractive milling technique using ExoCad software. Figure (2)




Fig 2: Digital scans to the typodont. a), b) prepared abutments buccally, c) prepared abutments occlusally, d) opposing mandibular arch.

A total of sixteen samples (eight in each group) were used in the present study. Group (I): included eight bridges (n=8) fabricated by CAD/CAM (control). Group (II): included eight bridges (n=8) fabricated by 3D printing (Intervention). The provisional bridges in the two group were numbered from 1 and ascending to 16, each sample was placed in an opaque sealed envelope, then were divided by (www.random.org) website into 2 equal groups in order to measure marginal fit and internal adaptation of two provisional bridges corresponding to each other randomly. External assessor was blinded (Single blind). Assessor received 16 sealed envelopes without knowing whether it belong to control or intervention group for randomization.

Measurements

Measurement of marginal gap was done by placing provisional bridges on the master die and fixed in place using special holding jig that help in holding each specimen throughout the process of gap evaluation. Measurements were taken in microns by a compatible personal computer using a fixed magnification of X40 attached with built in camera that was also connected to US digital microscope. Each sample

was photographed. Shots of the margins taken for each specimen. Shots were used in order to make morphometric measurements, Three equidistant landmarks along the cervical part of each specimen. Figure (3)




Fig 3: Photograph of 3D printed FDPs on the typodont abutment under digital microscope

Measurement of internal fit was done by a replica technique. On the intaglio of provisional bridges light body silicone (Elite, Zhermach, Italy) was added and seated into prepared abutments on the cast by finger pressure. Once light body silicone complete set, provisional bridge was removed. heavy-body silicone was used to stabilize the light-body silicone. Replicas were carefully sectioned into four equal segments by a razor blade (n°. 15c). To measure internal fit Two opposite sections were used. Figure (4)




Fig 4: Two opposite sections of the same replica.

Four regions were evaluated on each sectioned replica (axial wall, finish line, incisal \ occluso-axial and incisal \ occlusal), giving eight internal measures for each replica. At 25 × magnification Using built in camera (U500 X Digital Microscope, China) and USB digital microscopy, measurements were taken in microns.

Results

The results were found using Graph Pad Instat (Graph Pad, Inc.) software for windows. A value of $P \leq 0.05$ was considered statistically significant. Continuous variables were translated as the mean and standard deviation. After homogeneity of variance and normal distribution of errors had been confirmed, one-way analysis of variance was performed followed by Tukey’s post-hoc test if showed significance. Student t-test was done for compared pairs. Two-way ANOVA test was done to detect effect of variables influencing mean value. Sample size (n =8) was large enough to detect large effect sizes for main effects and pair-wise comparisons, with the satisfactory level of power set at 80% and a 95% confidence level.

Vertical Marginal fit measurement

It was analyzed that Milled group recorded statistically non-significant higher vertical marginal gap mean value ($39.38 \pm 7.57 \mu\text{m}$) than 3D Printed group ($33.23 \pm 7.98 \mu\text{m}$) as proven with two - way ANOVA test ($p = 0.0926 > 0.05$). Table (1) Figure (5).

Table 1: Comparison of total vertical marginal gap results (Mean values±SDs) as function of material group.

Variable		Vertical marginal gap		
		Mean±SD	95% CI	
			Low	High
Material	3D Printed	33.23±7.98	27.7	38.76
	Milled	39.38±7.57	34.14	44.62
Statistics	P value	0.0926 ns		

ns; non-significant ($p > 0.05$)




Fig 5: Column chart comparing total vertical marginal gap mean values as function of material group

Internal fit measurement

It was discovered that 3D Printed group was statistically non-significant greater internal gap mean value ($71.07 \pm 8.09 \mu\text{m}$) than Milled group ($65.73 \pm 6.1 \mu\text{m}$) as proven with two-way ANOVA test ($p = 0.3279 > 0.05$). Table (2) Figure (6)




Fig 6: Column chart comparing total internal gap mean values as function of material group.

Discussion

Provisional bridges are seated on the typodont using holding jig to hold the bridge with standardized force. all the procedure were done without cementation to shut out the effect of cementation technique disparity.

Vertical marginal gap ranges from 10 to 160 μm to be clinically acceptable. Moldoven *et al.*, 2006 stated that 100 μm as good marginal misfit and 200 to 300 μm as acceptable misfit. Recently most investigators recommend conclusion stated by McLean and von Fraunhofer that to be clinically accepted maximum Vertical marginal gap would be 120 μm The marginal and the internal discrepancy values of the digitalised FDP created using an IO scanner were, on average, lower than those measured by extra oral dental scanners used in laboratories, (Ozal Ç and Ulusoy M, 2021) (Dupagne *et al.*, 2023)

In the present study, it was discovered that the vertical marginal gap mean value recorded to the milled bridges was ($39.38 \pm 7.57 \mu\text{m}$) while the 3D printed group mean value was ($33.23 \pm 7.98 \mu\text{m}$) which is within clinically accepted range. The 3D printed provisional bridges showed greater marginal adaptation and fewer marginal inconsistency It's possible that the milling machine and end mill error have some sort of restriction on them. (Choi *et al.*, 2019)^[4].

This conclusion is in agreement with several other research, Park *et al.* in 2016 indicated that the marginal discrepancy value of implant interim restorations in the 3D printed group was superior than that in the subtractive milling approach

(CAM). He linked positive or negative error to milling bur diameter that should be considered. Although the 3D printing process was extremely precise and quick to carry out, the end result was an equally moulded layer that had a uniform face.

Elfar *et al.* in 2018 [7] examined marginal adequacy of restoration that manufactured by subtractive milled wax patterns & 3D printed wax patterns. It was discovered that the subtractive group had the largest mean marginal gap value (42.18 ± 1.44), while the 3D printed group had the lowest value (40.33 ± 0.77). These findings were discovered through research. According to him, the utilisation of 3D printing produced the most exact marginal adaption due to the creation of a sequences of cross - sectional layers. After that, every layer is printed on above the other in order to build three-dimensional body with less of a marginal difference and enough compensation for the polymerization shrinkage.

Alharbi *et al.* in 2018 shown that the marginal gap of the additive bridgess are much inferior to the subtractive counterparts. It is possible that errors brought on by the tolerance of milling burs are to blame for the marginal fit of milled restorations being inferior. If this is the case, then any surface details that are lesser than milling bur diameter will be excessively milled, which result in a improper fit of the bridges. Opposing results obtained by Alshalan *et al.* in 2019, They found that whether the software design of CAD/CAM system was subtractive or additive it had an effect on the marginal adaption of the restoration. In addition, despite the numerous benefits of the additive technique, its marginal fit is inferior to that of the subtractive technique. Due to the shrinkage that occurred throughout the building process, post-curing, and the minimum thickness of the layers.

The retention depends mostly on the correct fit of provisional restoration, The evaluation of internal fit is one of the most clinical significant outcomes. Failure on internal fit (misfit) is ascribed to inadequate retention. Using silicon replica technique the internal fit was measured and magnification by stereomicroscope as it is a nondestructive, accurate, least distortion methodology for evaluation of internal fit that method was employed in agreement with Park *et al.*, in 2015 Bhaskaran *et al.*, in 2013 [3] found that the suitable internal gap ranges between 81 and 136 μm .

The provisional restoration was loaded with low viscosity light body silicone then was seated over the tyodont by finger pressure till setting as suggested by Seok-Joon, and Jin-Hyun in 2016. After that, a heavy body silicon substance was filledn order to stabilize the lighter material, which made it simpler to manipulate and segment in the following steps. (Tamac *et al.*, 2015) [18]

In the present investigation, 3D printed group had higher internal gap mean value ($71.07 \pm 8.09 \mu\text{m}$) than milled group mean value ($65.73 \pm 6.1 \mu\text{m}$), It is possible that this was caused by software mistakes that occurred during the process of translating the STL file into printable format (transferring it into layers), which resulted in an unsatisfactory presentation of the interior fit.

This result is in agreement with the findings of several other studies, including one conducted by Kim *et al.* in 2014 which the researchers compared the internal fit of a three-unit metal framework of fixed dental prostheses that was manufactured by subtractive and additive manufacturing techniques. They measured axial wall, and occlusal gaps. The results for the SM group were $87.20 \pm 38.95 \mu\text{m}$ and $138.34 \pm 44.15 \mu\text{m}$. In contrast, the results of the AM group came in at $103.44 \pm 39.99 \mu\text{m}$ and $238.16 \pm 86.72 \mu\text{m}$, respectively. It was determined that this was due to a problem with the software.

According to Kang *et al.* in 2018 [10], the internal fit of restorations created by 3D printing was inferior to that which was produced using subtractive technology. Furthermore, the milling group demonstrated highly reproducible data in comparison to that which was shown by the 3D printing group. During creation of the 3D printed group, there may have been an occurrence of random mistake brought on by the light diffraction phenomenon. This is the theory for the lower reproducibility of the latter group. As a result of the development of layers, the surface of the 3D printing revealed an inaccurate representation. Also, the restorations that were made by 3D printing support originally eliminated by utilising diamond disc, the components meticulously finished and polished by a rubber bur. Despite this, there were still remnants of the support, which is one of the drawbacks associated with the 3D printing technology.

The results were in disagreement with Mai *et al* in 2020 [12], who observed that the milling group showed considerably higher internal gap in the occlusal region, than 3D printed polyjet group. Crowns made by the PolyJet group have a higher degree of precision in the occlusal region. This result may be related to a different underlying fabrication technique.

Conclusion

Within the qlimitationsq of the present study, the following points could be concluded

- Provisional bridges fabricated by the two different methods of manufacturing (milling and 3D printing) The vertical marginal gap values were within the acceptable clinical range of one hundred twenty microns.
- The 3D printed provisional restorations presented equivalent marginal fit to that of the milled ones.

Funding: This research received no external funding.

Conflicts of Interest: The authors declare no conflict of interest.

Reference

1. Abduo J, Lyons K, Bennamoun M. Trends in computer-aided manufacturing in prosthodontics: a review of the available streams. International journal of dentistry, 2014.
2. Al-shalan A, Arwa A, Aldin D, Bin-Abbas S, Al-shmlani M, Asif Z. Marginal and internal fit of CAD CAM system: A literature review. The Saudi Dental Journal, 2019, 31, S21.
3. Bhaskaran E, Azhagarasan NS, Miglani S, Ilango T, Krishna GP, *et al.* Comparative evaluation of marginal and internal gap of Co-Cr copings fabricated from conventional wax pattern, 3D printed resin pattern and DMLS tech: an *in vitro* study. The Journal of Indian Prosthodontic Society. 2013;13:189-195.
4. Choi JW, Ahn JJ, Son K, Huh JB. Three-dimensional evaluation on accuracy of conventional and milled gypsum models and 3D printed photopolymer models. Materials. 2019;12(21):3499.
5. Dupagne L, Mawussi B, Tapie L, Lebon N. Comparison of the measurement error of optical impressions obtained with four intraoral and one extra-oral dental scanners of post and core preparations. Heliyon, 2023, 9(2).
6. Dureja I, Yadav B, Malhotra P, Dabas N, Bhargava A, Pahwa R. A comparative evaluation of vertical marginal fit of provisional crowns fabricated by computer-aided design/computer-aided manufacturing technique and direct (intraoral technique) and flexural strength of the

- materials: An *in vitro* study. The Journal of the Indian Prosthodontic Society. 2018;18(4):314.
7. Elfar M, Korsel A, Kamel M. Marginal fit of heat pressed lithium disilicate crowns fabricated by three-dimensional printed and subtractive CAD/CAM wax patterns. Tanta Dental Journal. 2018;15(4):199.
 8. Ha SJ, Cho JH. Comparison of the fit accuracy of zirconia-based prostheses generated by two CAD/CAM systems. The journal of advanced prosthodontics. 2016;8(6):439-448.
 9. Jeong YG, Lee WS, Lee KB. Accuracy evaluation of dental models manufactured by CAD/CAM milling method and 3D printing method. The journal of advanced prosthodontics. 2018;10(3):245-251.
 10. Kang SY, Park JH, Kim JH, Kim WC. Accuracy of provisional crowns made using stereolithography apparatus and subtractive technique. The Journal of Advanced Prosthodontics. 2018;10(5):354.
 11. Kim KB, Kim JH, Kim WC, Kim JH. *In vitro* evaluation of marginal and internal adaptation of three-unit fixed dental prostheses produced by stereolithography. Dental Materials Journal. 2014;33(4):504-509.
 12. Mai HY, Lee WK, Kwon TG, Lee DH. Reliability of digital measurement methods on the marginal fit of fixed prostheses: A systematic review and meta-analysis of *in vitro* studies. The Journal of prosthetic dentistry. 2020;124(3):350-e1.
 13. McLean JW. The estimation of cement film thickness by an *in vivo* technique. British dental journal. 1971;131(3):107-111.
 14. Moldovan O, Rudolph H, Quaas S, Bornemann G, Luthardt RG. Internal and external fit of CAM-made zirconia bridge frameworks-a pilotstudy. *Deutsche Zahnärztliche Zeitschrift*. 2006;61(1):38.
 15. Özal Ç, Ulusoy M. *In-vitro* evaluation of marginal and internal fit of 3-unit monolithic zirconia restorations fabricated using digital scanning technologies. The Journal of Advanced Prosthodontics. 2021;13(6):373.
 16. Park JK, Lee WS, Kim HY, Kim WC, Kim JH. Accuracy evaluation of metal copings fabricated by computer-aided milling and direct metal laser sintering systems. The journal of advanced prosthodontics. 2015;7(2):122-128.
 17. Park JY, Lee JJ, Bae SY, Kim JH, Kim WC. *In vitro* assessment of the marginal and internal fits of interim implant restorations fabricated with different methods. The Journal of prosthetic dentistry. 2016;116(4):536-542.
 18. Tamac E, Toksavul S, Toman M. Clinical marginal and internal adaptation of CAD/CAM milling, laser sintering, and cast metal ceramic crowns. The Journal of prosthetic dentistry. 2014;112(4):909-913.